



Mastercool[®]
"World Class Quality"

CERTIFICATE OF CALIBRATION

Applicant:

Instrument:

Electronic Leak Detector

Serial Number:

Calibration Number:

Calibration Method:

Three tests done at high sensitivity setting with specified leak levels of 3 g, 5 g and 10 g/year. Results shown are pass/fail at a reaction time of +/- 0,5 sec. This leak detector has been tested against a calibrated supply source from LACO Technologies Leak Standard, Model CM5HX-4305A. All reference standards used in the calibration of the LACO technologies leak source are traceable to the National Institute Standards and Technology (NIST)

Environmental Condition:

Temperature: --°C

Calibration Results		
Measurement	Sensitivity Level	Results
3 grams	Highest Level	
5 grams	Highest Level	
10 grams	Highest Level	

Date of Test:

Certificate Validity:

Approved By: Kris Luyckx

Date:

MASTERCOOL INC.

One Aspen Drive Randolph, NJ 07869 USA • Phone: 973-252-9119 • Fax: 973-252-2455
Laagstraat 19 • 9140 Temse, Belgium • Telefon: +32 (0) 3 777 28 48 • Telefax: +32 (0) 3 777 40 62